

Product group: **661** Product number: **778416**

Intestinal Enterococci are bacteria that can be used as a marker to indicate fecal contamination of Potable Water or Ballast Water.

Product information

The test is designed to detect Enterococcus species in drinking water samples within 24 hours. When the substrate is metabolised by enterococci the sample turns from blue to green to indicate a positive detection. Any change from the original colour to green is considered a positive result. No ultraviolet light source is required.



Test Paramater: Bacteria - Intestinal Enterococci

Used with:

- 1. POTABLE WATER TEST KIT (778419)
- 2. BALLAST CHECK 2 TEST KIT (777848)

Features

- Simple presence/absence test
- Easy to carry out on board ship

Benefits

- Portable, easy to use for testing of potable water and ballast water
- Optional: For Ballast water, dilute the sample by 1 part sample and 9 part deionised water (10 ml sample + 90 ml demin water = 100 ml sample)

Specification

General

Invent Hazard Material (IMO/IU) classification	NA
--	----

Documents

[SDoC and MD for IHM](#)

Related products

Is accessory to

778419
POTABLE WATER TEST KIT
777848
BALLAST CHECK 2 TEST KIT

Is frequently bought together with

777722
LEGIONELLA TEST KIT...
778413
BACTERIA PLATES (40 PCS)
778414
DILUTION POTS (40 PCS)
778421
CHECKIT FREE CHLORINE DISC
777932
E.COLI/COLIFORM (25 TESTS)

This page is printed from:

<https://www.wilhelmsen.com/product-catalogue/products/marine-chemicals/test-kits-and-reagents/ballast-w ater-testing/intestinal-enterococci-kitpk20/>